



DATASHEET : EN NORMS APPROVED & TESTED BY ACCREDITED LABS.

EN-tests	Norm year	PT-group	Test strains	Type	Contact time Minutes	Temp. Celsius	Solution %	Soiling g/L Albumin	Logs >=	Tested by Accredited lab.
1276	2019	2.4	Staphylococcus aureus ATCC 6538 in clean conditions	Bactericidal	1	20	50	0.3g/L	5	Melbec
			Enterococcus hirae ATCC 10541 in clean and dirty conditions		1	20	50	0.3g/L	5	Melbec
			Escherichia coli ATCC 10536 in clean conditions		1	20	50	0.3g/L	5	Melbec
			Pseudomonas aeruginosa ATCC 15442 in clean conditions		1	20	50	0.3g/L	5	Melbec
1650	2013	2.4	Candida albicans ATCC 10231	Yeasticidal	5	20	80	0.3g/L	4	HygCen
13624	2013	2	Candida albicans ATCC 10231	Yeasticidal	3	20	80	0.3g/L	4	Melbec
13697	2019	2.4	Staphylococcus aureus ATCC 6538 in clean and dirty conditions	Bactericidal	1	20	50	0.3g/L	4	Melbec
			Enterococcus hirae ATCC 10541 in clean and dirty conditions		1	20	50	0.3g/L	4	Melbec
			Escherichia coli ATCC 10536 in clean and dirty conditions		1	20	50	0.3g/L	4	Melbec
			Pseudomonas aeruginosa ATCC 15442 in clean and dirty conditions		1	20	50	0.3g/L	4	Melbec
			Candida albicans ATCC 10231 in clean and dirty conditions	Yeasticidal	5	20	50	0.3g/L	3	Melbec
13704	2002	2.4	Clostridium difficile ATCC 9689	Sporicidal	60	20	50	0.3g/L	3	HygCen
13727	2014	2	Staphylococcus aureus ATCC 6538	Bactericidal	1	20	50	0.3g/L	5	HygCen
			Enterococcus hirae ATCC 10541		1	20	50	0.3g/L	5	HygCen
			Pseudomonas aeruginosa ATCC 15442		1	20	50	0.3g/L	5	HygCen
			Mycobacterium terrae ATCC 15755		10	20	97	0.3g/L	5	HygCen
14348	2005	2.4	Mycobacterium avium ATCC 15769	Mycobactericidal	10	20	97	0.3g/L	5	HygCen
14476	2013	2.4	Poliovirus Type 1 strain LSc-2ab	Virucidal	2	20	50	0.3g/L	4	HygCen
			Murine Norovirus strain 599 Berlin		2	20	80	0.3g/L	4	HygCen
			Adenovirus Type 5 strain Adenoid 75, ATCC VR 5		2	20	50	0.3g/L	6	HygCen
			Staphylococcus aureus ATCC 6538		1	20	80	0.3g/L	5	HygCen
14561	2006	2	Enterococcus hirae ATCC 10541	Bactericidal	1	20	80	0.3g/L	5	HygCen
			Pseudomonas aeruginosa ATCC 15422		1	20	50	0.3g/L	5	HygCen
			Candida albicans ATCC 10231		5	20	100	0.3g/L	4	HygCen
14562	2006	2	Candida albicans ATCC 10231	Yeasticidal	5	20	100	0.3g/L	4	HygCen
14563	2009	2	Mycobacterium terrae ATCC 15755	Mycobactericidal	10	20	80	0.3g/L	5	HygCen
			Mycobacterium avium ATCC 15769		10	20	80	0.3g/L	5	HygCen
16615	2015	2.4	Staphylococcus aureus ATCC 6538	Bactericidal	1	23 / 23.5	100	0.3g/L	5	HygCen
			Enterococcus hirae ATCC 10541		1	23 / 23.5	100	0.3g/L	5	HygCen
			Pseudomonas aeruginosa ATCC 15442		1	23 / 23.5	100	0.3g/L	5	HygCen
			Candida albicans ATCC 10231	Yeasticidal	1	20	100	0.3g/L	4	Melbec
			Candida auris BSM 21092	Yeasticidal	1	20	100	0.3g/L	4	Melbec
16777	2018	2.4	Vaccinia virus VR-1508	Virucidal 24-hour residual	2	20	50	0.3g/L	4	Melbec
PAS2424	NA	2.4	Staphylococcus aureus ATCC 6538	Bactericidal 24-hour residual	5	20	50	0.3g/L	4	Melbec
			Enterococcus hirae ATCC 10541		5	20	50	0.3g/L	4	Melbec
			Escherichia coli ATCC 10536		5	20	50	0.3g/L	4	Melbec
			Pseudomonas aeruginosa ATCC 15442		5	20	50	0.3g/L	4	Melbec
			Candida albicans ATCC 10231	Yeasticidal 24-hour residual	15	20	50	0.3g/L	3	Melbec
1656	2010	3	Staphylococcus aureus ATCC 6538	Bactericidal	30	10	50	0.3g/L	5	HygCen
			Enterococcus hirae ATCC 10541		30	10	10	0.3g/L	5	HygCen
			Proteus hauseri (formerly vulgaris) ATCC 13315		30	10	10	0.3g/L	5	HygCen
			Pseudomonas aeruginosa ATCC 15442		30	10	50	0.3g/L	5	HygCen
1657	2014	3	Candida albicans ATCC 10231	Yeasticidal	30	10	97	0.3g/L	4	HygCen
14349	2013	3	Staphylococcus aureus ATCC 6538	Bactericidal	30	10	50	0.3g/L	4	HygCen
			Enterococcus hirae ATCC 10541		30	10	50	0.3g/L	4	HygCen
			Pseudomonas aeruginosa ATCC 15442		30	10	50	0.3g/L	4	HygCen
			Proteus hauseri (ehemals / formerly P. vulgaris) ATCC 13315		30	10	10	0.3g/L	4	HygCen
14675	2015	3	Bovine enterovirus Type 1 (ECBO) ATCC VR-248	Virucidal	30	10	80	0.3g/L	4	HygCen
16438	2014	3	Candida albicans ATCC 10231	Yeasticidal	30	10	50	0.3g/L	3	HygCen